天津市津腾实验设备有限公司
Tianjin JinTeng Experiment Equipment Co., Ltd

Factory Inspection Report

Report No.：JT-FM-8233-004 Page 1/1

Product Name	Membrane Filter	Report Date	Sep,4th,2026
Material	GF Membrane	Quantity	5000pcs
Diameter& Pore Size	47mm--0.45um	Inspection Date	Sep,2nd,2026

Inspection Item	Inspection Standard	Sampling Quantity	Result
Appearance	The surface shall be flat and clean, dry with uniform color, free from impurity contamination, folds and mechanical damage.	200pcs	Pass
Thickness（μm）	353	70pcs	Pass
Flow Rate(s)	4.1s@-5	70pcs	Pass
Bubble Point(kPa)	17.6	70pcs	Pass
Diameter(mm)	47	70pcs	Pass
Conclusion	Pass the inspection and be allowed to leave the factory		

Dept: Quality Department Reviewer：范志明 Analyst：何建波